

Deluxe Guard Rail Pads with padding help protect patients from injury by helping close some of the gaps along the side rail of the mattress and by cushioning the side rails.

Each Deluxe Guard Rail Pad attaches with hook and loop straps and is made of durable, firm density foam and covered in soft, water repellent vinyl. Meets CA #117, Section E Flame Retardancy Standards.

<b>AL-58230</b>	72"L x 14"H (183 cm x 36 cm)
<b>5605W</b>	72"L x 14"H (183 cm x 36 cm) with window
<b>5606</b>	60"L x 14"H (152 cm x 36 cm)
<b>5607</b>	43"L x 14"H (109 cm x 36 cm)

### Monitoring

**WARNING** Deluxe Guard Rail Pads will not prevent a determined patient from exiting the bed, going over or attempting to go between or through side rail gaps.

**ALWAYS** adjust straps when articulating the bed to help prevent the product from coming loose or being damaged.

**ALWAYS** follow your facility's policies and procedures for patient assessment, monitoring and rehabilitation.

### Application Instructions:

1. Inspect product prior to use. DO NOT use soiled or damaged products.
2. Raise the side rails to the "UP" position. Separate the hook and loop straps.
3. Position the side rail pad on the inside of the side rail with the straps facing out away from the bed. The top strap section should hang over the side rail.
4. Pull each strap tight around the vertical side rail bar and secure (model 5605W only).
5. Check for proper fit. Modified beds may not be compatible with Deluxe Guard Rail Pads.




### Cleaning Instructions:

Wipe Clean with soap and water or disinfectant.

DO NOT use cleaners that contain Phenol or Benzyl.

**WIPE CLEAN**



### Disposal

**WARNING** Properly dispose of the product per facility's policy for BIOHAZARDOUS materials.

### Storage And Handling

This device is designed for use in normal indoor environments. This device may be stored in ambient warehouse temperatures at normal humidity levels. Avoid excess moisture or high humidity that may damage product materials.