

# AL-78569 Corridor Dome Light Specification Sheet



# ARCHITECTS' AND ENGINEERS' SPECIFICATION

Corridor light shall be ALCO AL-78569 Corridor Dome Light or approved equal. The corridor light shall be sectional in design. Lights shall include LEDs to provide amber, green, red and white colored annunciation through a translucent plastic lens. All lenses shall be mounted on a flame-retardant (UL® 94HB) plastic base panel. The four colored LEDs shall indicate patient call priority and staff presence.

The AL-78569 shall be UL® Listed.

#### **FEATURES**

ALCO's AL-78569 Corridor Dome Light provides visual indication of calls originating from patient, staff, duty, bath, emergency, presence and code call stations. Patient call priority and staff presence are indicated by amber, green, red and white colored LEDs. When used with any room station or Staff Station, the AL-78569 indicates all calls originated from that room. When used with the Multipurpose Station or the Duty Station, the AL-78569 may be used as a zone light to indicate all calls from the zones to which those stations are assigned.

- UL® 1069 Listed
- Long lasting LEDs
- · No bulb replacement required
- Plug-in connector supplied
- Attractive flame-retardant plastic panel

## **SPECIFICATIONS**

Dimensions: Height: 4.8" (122 mm)

Width: 5.0" (127 mm)
Depth: 2.755" (70 mm)

LEDs: 2V 20mA per LED

Power: 24 VDC

Finish: Base is flame-retardant (UL® 94HB)

plastic. Lens is flame-retardant translucent polycarbonate plastic.

Housing: Fits standard one- or two-gang box with

one- or two-gang ring.

## REQUIRED COMPONENTS

Tek-CARE®NC300II Nurse Call System